OHIO DEPARTMENT OF HEALTH

ALCOHOL AND DRUG TESTING PROGRAM

APPROVAL OF INSTRUMENT CHECK SOLUTION, OAC 3701-53-04

Lot or Batch # 11020 Target Value 0.101 g/210L

This instrument check solution is to be used in a simulator for the instrument check of approved evidential breath testing instruments, as defined in OAC 3701-53-02.

This instrument check solution contains **1.225 mg/mL** ethyl alcohol in purified water. When used according to instrument check forms, it will produce a result at or within ±0.005 g/210L of the target value in the approved breath testing instrument, when that instrument is in proper working condition.

This instrument check solution shall not be used more than three months after its date of first use or more than one year after its date of manufacture, **January 10, 2011.** Whichever comes first.

This solution expires January 10, 2012

Approved this 15th day of February, 2011

Director of Health

© Goes 461½

LITHO. IN U.S.A.