

**3701:1-48-06 Limits on external radiation levels from storage containers
and source changers.**

The maximum exposure rate limits for storage containers and source changers are two millisievert (two hundred millirem) per hour at any exterior surface, and 0.1 millisievert (ten millirem) per hour at one meter from any exterior surface with the sealed source in the shielded position.

Five Year Review (FYR) Dates: 04/04/2016 and 04/01/2021

CERTIFIED ELECTRONICALLY

Certification

04/04/2016

Date

Promulgated Under:	119.03
Statutory Authority:	3748.04
Rule Amplifies:	3748.06, 3748.07
Prior Effective Dates:	10/20/2002, 3/22/07